


Traceable Certificate Number: 4067323
Contractor: BASTROP SCALE COMPANY
 PO BOX 2100
 BASTROP, TX 78602-9100

Purchase Order Number: 260508KB01
Client: BASTROP SCALE COMPANY
 192 HARMON RD
 BASTROP, TX 78602

Date Received: 28 May 2026
Date Calibrated: 29 May 2026 to 03 Jun 2026
Recalibration Date: 29 May 2027
NIST Certificate Number: 684/O-0000078300

If there are two NIST numbers, one or both may apply

Calibrated By: 05, 38
Procedure: WI05-0095 Rev. G
Condition of Weights: Acceptable for Calibration
Description of Weights: 1 mg to 100 g Polished Weight Set, ASTM Class 1, S/N AP-1M89
Comments:



Key Notes

- Finish ✱ Indicates the weight does not meet the finish requirements
- Material ⊕ Indicates the weight does not meet the material requirements
- New Wt ◇ Indicates new weight
- Missing Wt ▲ Indicates replaced missing weight with new weight
- Damaged Wt ✕ Indicates replaced damaged weight
- Replaced OOT ★ Indicates replaced out of tolerance weight
- OOT ⊠ Indicates correction plus or minus Uncertainty greater than or equal to MPE
- Magnetic Wt ★★ Indicates replaced magnetic weight
- Design ⊗ Indicates the weight does not meet the design or shape requirements
- Repainted 🗑 Indicates the weight was repainted after As Found obtained
- Other ⊕ See comments above

Cleaning Levels

- A Dusted with brush or cloth
- B Spot cleaned with ethyl alcohol
- C Full surface cleaned with ethyl alcohol
- D Spot cleaned with non-alcohol solvent followed by ethyl alcohol
- E Full surface cleaned with non-alcohol solvent followed by ethyl alcohol
- F No cleaning performed

Material Abbreviations

- | | |
|--------------------|------------------|
| AL Aluminum | TA Tantalum |
| SS Stainless Steel | BR Brass |
| CI Cast Iron | PL Platinum |
| IR Iron | NS Nickel Silver |
| MS Mild Steel | OR Other/Unknown |

Check with your local state agency for certification of compliance on Legal-for-Trade items. The weight accuracy class is referenced in the Description of Weights. Unless otherwise noted, the weights calibrated meet the requirements of the accuracy class. Results relate only to weights calibrated. The Surface Finishes of weights are evaluated visually. Weights are screened for magnetism using work instruction WI05-0035 when they are new, when requested by the customer or when weights are suspected of not meeting specifications. Density if measured is measured using OIML R111-1 (2004) method A2. Conventional Mass is reported based on a reference density of 8.0 g/cm³. The Uncertainty of Measurement is included in the determination of Maximum Permissible Error (MPE) Pass/Fail Criteria. The specifications for Maximum Permissible Error (MPE) can be found in NIST Handbook 105-1 (2019), NIST Handbook 105-1 (1990), ASTM E617-23 or OIML R111-1 (2004), manufacturer specifications or customer specifications.

The Uncertainty assigned to the Conventional Mass values are the result of the root-sum-square of the type A and type B components, calculated in accordance with NIST SOP 29 and the Guide to the expression of uncertainty in measurement, with coverage factor (k=2), to express the expanded uncertainty with an approximate 95.45% confidence level. This report is not to be used to claim product certification, approval, or endorsement by NVLAP, NIST, A2LA, or any government agency. **This document and all data within, shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems.**

Dan Demers
 Dan Demers, Metrologist



Prepared By:
Rice Lake Weighing Systems ● PN 64784 ● 6/25
 230 West Coleman Street ● Rice Lake, WI 54868 ● USA
 TEL: 715-234-9171 ● FAX: 715-234-6967

Definitions: http://certs.ricelake.com/certs/0354_Term_Cert_Weight_Cal_Rev1.pdf
 Page 1 of 2

03 Jun 2026

Issued Date:



Certificate of Weight Calibration



ISO/IEC 17025:2017 & ANSI/INCSL-Z540-1-1994 ACCREDITED
 Temperature Range: 21.92 °C to 22.13 °C
 Pressure Range: 730.79 mmHg to 735.28 mmHg
 Relative Humidity Range: 51 % to 56 %

Traceable Certificate Number: 4067323
 Client: BASTROP SCALE COMPANY
 Date Calibrated: 29 May 2026 to 03 Jun 2026

As Left Data (As Found Data is undifferentiated from As Left Data unless listed in As Found Data table)

Nominal Value	Unique ID	True Mass (Same UOM as Nom.)	True Mass Corr. (mg)	Conv. Mass (Same UOM as Nom.)	Conv. Mass Corr. (mg)	(k=2) Unc. (± mg)	MPE (± mg)	MPE Pass (Y=Pass, N=Fail)	Assumed Density (g/cm ³)	Assumed Material	Const. Type	Balance Used	Reference Standard Set Used	Air Density (mg/cm ³)	Clean Level
1 mg	AP-1M89	1.00445	0.00445	1.00444	0.00444	0.00098	0.010	Y	7.95	SS	I	503Q	T535Q	1.1517	A
2 mg	AP-1M89	2.00334	0.00334	2.00334	0.00334	0.00075	0.010	Y	7.95	SS	I	503Q	T535Q	1.1512	A
2 mg	AP-1M89	2.00489	0.00489	2.00489	0.00489	0.00075	0.010	Y	7.95	SS	I	503Q	T535Q	1.1511	A
5 mg	AP-1M89	5.00147	0.00147	5.00147	0.00147	0.00094	0.010	Y	7.95	SS	I	503Q	T535Q	1.1514	A
10 mg	AP-1M89	9.9942	-0.0058	9.9941	-0.0059	0.0011	0.010	Y	7.95	SS	I	503Q	T535Q	1.1517	A
20 mg	AP-1M89	20.00145	0.00145	20.00143	0.00143	0.00098	0.010	Y	7.95	SS	I	503Q	T535Q	1.1516	A
20 mg	AP-1M89	19.99468	-0.00532	19.99466	-0.00534	0.00098	0.010	Y	7.95	SS	I	503Q	T535Q	1.1516	A
50 mg	AP-1M89	50.00119	0.00119	50.00114	0.00114	0.00072	0.010	Y	7.95	SS	I	2022Q	L595Q	1.1438	A
100 mg	AP-1M89	99.99599	-0.00401	99.99589	-0.00411	0.00070	0.010	Y	7.95	SS	I	2022Q	L595Q	1.1438	A
200 mg	AP-1M89	199.99749	-0.00251	199.99730	-0.00270	0.00060	0.010	Y	7.95	SS	I	2022Q	L595Q	1.1438	A
200 mg	AP-1M89	199.99989	-0.00011	199.99970	-0.00030	0.00060	0.010	Y	7.95	SS	I	2022Q	L595Q	1.1438	A
500 mg	AP-1M89	500.00547	0.00547	500.00500	0.00500	0.00078	0.010	Y	7.95	SS	I	2022Q	L595Q	1.1438	A
1 g	AP-1M89	1.0000017	0.0017	1.0000007	0.0007	0.0011	0.034	Y	7.95	SS	I	2022Q	L595Q	1.1438	A
2 g	AP-1M89	1.9999847	-0.0153	1.9999828	-0.0172	0.0013	0.034	Y	7.95	SS	II	2022Q	L595Q	1.1438	A
2 g	AP-1M89	1.9999961	-0.0039	1.9999942	-0.0058	0.0013	0.034	Y	7.95	SS	II	2022Q	L595Q	1.1438	A
5 g	AP-1M89	5.0000172	0.0172	5.0000125	0.0125	0.0029	0.034	Y	7.95	SS	II	2022Q	L595Q	1.1438	A
10 g	AP-1M89	10.0000005	0.0005	9.9999911	-0.0089	0.0057	0.050	Y	7.95	SS	II	2060Q	U520Q	1.1453	A
20 g	AP-1M89	20.0000202	0.0202	20.0000014	0.0014	0.0054	0.074	Y	7.95	SS	II	2060Q	U520Q	1.1453	A
20 g	AP-1M89	20.0000412	0.0412	20.0000224	0.0224	0.0054	0.074	Y	7.95	SS	II	2060Q	U520Q	1.1453	A
50 g	AP-1M89	50.0000068	0.068	50.0000021	0.021	0.012	0.12	Y	7.95	SS	II	2060Q	U520Q	1.1453	A
100 g	AP-1M89	100.0000018	0.018	99.9999924	-0.076	0.023	0.25	Y	7.95	SS	II	2060Q	U520Q	1.1453	A