



TEXAS DEPARTMENT OF AGRICULTURE

COMMISSIONER SID MILLER

Metrology Laboratory - 1258 CR 226 / P.O. Box 1518 - Giddings, Texas 78942

Phone: (979) 542-3231 - Fax: (888) 205-7741

Page1 of 3

CALIBRATION CERTIFICATE

Calibration Number

G-0000010007

For

16 - 1000 lb Weights

13 - 50 lb Weights

2 - 25 lb Weights

Submitted By

Bastrop Scale Company Incorporated

PO Drawer 2100

Bastrop, Texas 78602

512-321-3443

The measurement results of the Texas Department of Agriculture, Giddings Metrology Laboratory are traceable to the International System of Units (SI) via the Consensus Value through the measurements at the National Institute of Standards and Technology (NIST) and are a part of comprehensive measurement assurance program for ensuring continuous accuracy and measurement traceability within the level of the uncertainty reported by this laboratory. The laboratory calibration number above is the unique report number to be used in referencing measurement traceability for artifacts identified in this certificate only. The data applies only to the artifacts identified in this certificate at the time of test. Calibration certificate shall not be reproduced, except in full, without written laboratory approval.

Calibration Date: 05/29/2026

Calibration Due: 05/31/2027

Issue Date: 05/29/2026

Average Temperature: 20.84 °C

Average Humidity: 52.01 %

Procedure: NISTIR 6969, SOP No. 8, Modified Substitution (Rev. 2019)

Mass Standards: Giddings Metrology Laboratory Mass Echelon III Standards

Received Date: 05/28/2026

Condition Received: Acceptable

For cast-iron weights 10 kg and larger, acceptance tolerances corresponding to ASTM Class 6 (numerically equivalent to NIST Class F) have been applied. Listing ASTM Class 6 in the tolerance class column reflects the tolerance criteria used for evaluation only and does not constitute assignment, verification, or reclassification of weight class. Only compliance with tolerance specifications were evaluated for items listed on this certificate (failing values are indicated in the table, if any.) The uncertainty of the measurement was taken into account when making this statement of compliance. The weights were not evaluated for conformance with technical requirements (design, construction, material, magnetism, density, surface finish and marking.) Tolerances were taken from NIST 105-1 (1990), ASTM E617 (2023) or OIML R111 (2004).

The combined standard uncertainty consists of both Type A and Type B components, including the standard uncertainty reported for the standard, the standard uncertainty for the measurement process, and a component of uncertainty to account for any observed deviations that have a significant effect on the calibration combined, using the root sum square method. Air buoyancy was considered negligible and was not included. The uncertainty does not include contribution due to magnetism or irregular conditions on the surface of the weights. The expanded uncertainty given is in compliance with BIPM JCGM 100:2008, Guide to the Expression of Uncertainty in Measurement (GUM), 2008 and follows NISTIR 6969, SOP 29 (2019), with a variable k (coverage factor) representing a 95.45 % confidence level.

Note:

A positive correction indicates that the weight is heavier than the stated nominal value.

A negative correction indicates that the weight is lighter than the stated nominal value.

Conversions:

milligram (mg) to kilogram (kg): $kg = mg / 1000000$

milligram (mg) to gram (g): $g = mg / 1000$

milligram (mg) to pound (lb): $lb = mg \times 0.000002204622621848776$

milligram (mg) to ounce (oz): $oz = mg \times 0.00003527396194958041$

This certificate must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

Lisa Corn

Manager for Metrology Laboratory
Agency Representative



Keri Schatte
Metrologist
Approved Signatory



TEXAS DEPARTMENT OF AGRICULTURE

Page 3 of 3

COMMISSIONER SID MILLER

Metrology Laboratory - 1258 CR 226 / P.O. Box 1518 - Giddings, Texas 78942

CALIBRATION CERTIFICATE

For

16 - 1000 lb Weights

13 - 50 lb Weights

2 - 25 lb Weights

Calibration Number

G-0000010007

Calibration Date

05/29/2026

Date Due

05/31/2027

Submitted by

Bastrop Scale Company Incorporated

PO Drawer 2100

Bastrop, Texas 78602

Average Temperature: 20.84 °C

Average Humidity: 52.01 %

SOP Used: NISTIR 6969, SOP No. 8, Modified Substitution

Observations:

The artifacts described below have been compared to the standards of the State of Texas and were found to have the following mass corrections:

Nominal Value	Serial	ID #	As Found Mass Correction (mg)	As Left Mass Correction (mg)	Expanded Uncertainty (mg)	k factor	Tolerance Class	Tolerance Status	Tolerance (mg)
1000 lb	36HK	-	-52700	1400	6200	2.009	ASTM 6	In Tolerance	45000
1000 lb	36HY	-	-39800	3100	6200	2.009	ASTM 6	In Tolerance	45000
1000 lb	3610	-	-51200	1400	6200	2.009	ASTM 6	In Tolerance	45000
1000 lb	3612	-	-49000	1900	6200	2.009	ASTM 6	In Tolerance	45000
1000 lb	3615	-	-26300	-26300	6200	2.009	ASTM 6	In Tolerance	45000
1000 lb	36HZ	-	-75900	2200	6200	2.009	ASTM 6	In Tolerance	45000
1000 lb	3611	-	-11900	-11900	6200	2.009	ASTM 6	In Tolerance	45000
1000 lb	3616	-	-73900	900	6200	2.009	ASTM 6	In Tolerance	45000
1000 lb	3619	-	-13200	-13200	6200	2.009	ASTM 6	In Tolerance	45000
1000 lb	3614	-	-14100	-14100	6200	2.009	ASTM 6	In Tolerance	45000
1000 lb	3613	-	1200	1200	6200	2.009	ASTM 6	In Tolerance	45000
1000 lb	3618	-	-26400	-26400	6200	2.009	ASTM 6	In Tolerance	45000
1000 lb	361A	-	-29100	-29100	6200	2.009	ASTM 6	In Tolerance	45000
1000 lb	361B	-	-63100	500	6200	2.009	ASTM 6	In Tolerance	45000
1000 lb	361C	-	-11200	-11200	6200	2.009	ASTM 6	In Tolerance	45000
1000 lb	3617	-	-30300	-30300	6200	2.009	ASTM 6	In Tolerance	45000
50 lb	BS6209	-	-10000	-10	300	2.002	ASTM 6	In Tolerance	2300
50 lb	5PPD	-	-5520	10	300	2.002	ASTM 6	In Tolerance	2300
50 lb	BS57	-	-15250	40	300	2.002	ASTM 6	In Tolerance	2300
50 lb	BS61A	-	-7200	10	300	2.002	ASTM 6	In Tolerance	2300
50 lb	5PPB	-	-11090	-20	300	2.002	ASTM 6	In Tolerance	2300
50 lb	BS78	-	-4680	-20	300	2.002	ASTM 6	In Tolerance	2300
50 lb	BS56A	-	-11100	-10	300	2.002	ASTM 6	In Tolerance	2300
50 lb	BS124	-	-13350	-30	300	2.002	ASTM 6	In Tolerance	2300
50 lb	BS24	-	-15870	-10	300	2.002	ASTM 6	In Tolerance	2300
50 lb	BS6199	-	-17500	-20	300	2.002	ASTM 6	In Tolerance	2300
50 lb	5PPC	-	-8170	20	300	2.002	ASTM 6	In Tolerance	2300
50 lb	5PPA	-	-340	-340	300	2.002	ASTM 6	In Tolerance	2300
50 lb	BS644	-	-7790	-10	300	2.002	ASTM 6	In Tolerance	2300
25 lb	BS65	-	-260	-260	140	2.006	ASTM 6	In Tolerance	1100

• denotes a weight that was adjusted per NISTIR 6969, SOP 8.

▲ denotes a weight that was rejected.

For cast-iron weights 10 kg and larger, acceptance tolerances corresponding to ASTM Class 6 (numerically equivalent to NIST Class F) have been applied. Listing ASTM Class 6 in the tolerance class column reflects the tolerance criteria used for evaluation only and does not constitute assignment, verification, or reclassification of weight class. Only compliance with tolerance specifications were evaluated for items listed on this certificate (failing values are indicated in the table, if any.) The uncertainty of the measurement was taken into account when making this statement of compliance. The weights were not evaluated for conformance with technical requirements (design, construction, material, magnetism, density, surface finish and marking.) Tolerances were taken from NIST 105-1 (1990), ASTM E617 (2023) or OIML R111 (2004).

The combined standard uncertainty consists of both Type A and Type B components, including the standard uncertainty reported for the standard, the standard uncertainty for the measurement process, and a component of uncertainty to account for any observed deviations that have a significant effect on the calibration combined, using the root sum square method. Air buoyancy was considered negligible and was not included. The uncertainty does not include contribution due to magnetism or irregular conditions on the surface of the weights. The expanded uncertainty given is in compliance with BIPM JCGM 100:2008, Guide to the Expression of Uncertainty in Measurement (GUM), 2008 and follows NISTIR 6969, SOP 29 (2019), with a variable k (coverage factor) representing a 95.45 % confidence level.



NVLAP Lab Code 600376-0

Lisa Corn
Manager for Metrology Laboratory
Agency Representative

Keri Schatte
Metrologist
Approved Signatory

