



Holloway

LIFTING & RIGGING SOLUTIONS

5833 ARMOUR DR. • HOUSTON, TEXAS 77026
PH: 713-674-5631 • www.hhtlfting.com
wirerope@hollowayhouston.com

Certificate of Calibration

Client: Bastrop Scale CO Inc.

Sales Order #: 1174958

Device calibrated: 300K Compression Load Cell
Type: Cabled
Manufacturer /model: Coti Inc./ CG26S5
Device serial #: 2011027002
Indicator type: GSE 250 Cabled Display
Indicator serial #: 2011027002

Date Calibrated: June 18, 2026
Verification Due: June 18, 2027

Temperature: 73°F
Humidity: 50%

Mode of calibration: Compression
Display units: LBF
Designated accuracy: ASTM E4-24
Tolerance (+/-): 1% of applied load
Received in tolerance: No
Adjustment made: Yes


Calibration factor: N/A

Reference True Load LBF	As Found Reading LBF	Run 1 Indication LBF	Percent Error of Reading	Meets ASTM E-4 Accuracy?	Run 2 Indication LBF	Percent Error of Reading	Meets ASTM E-4 Accuracy?	Repeatability as a percent
0	0	0			0			
30,000	29,700	30050	0.17	YES	30,100	0.33	YES	0.17
120,000	118,450	120200	0.17	YES	120,200	0.17	YES	0.00
180,000	177,550	180200	0.11	YES	180,200	0.11	YES	0.00
240,000	236,450	240000	0.00	YES	240,000	0.00	YES	0.00
300,000	295,200	299650	0.12	YES	299,650	0.12	YES	0.00


This device meets the accuracy standards of ASTM E4-24 as indicated above. Calibrations are performed in accordance with HHI specification #310001 Rev A; and all HHI reference standards used for calibration are traceable through NIST to the International System of Units (SI). This document complies with the requirements listed in ANSI/NCSL Z540.3-2006.

HHI reference standard(s):

Serial	Calibrated	Date Due	Class A loading range	Verified by	MFG by	Temperature
B263412	12/29/2025	12/29/2026	17,500 - 350,000 LBF	NSTL	Coti Inc.	23C

Calibration By: 

Date: 6/18/26

Approved By: 

Date: 6-18-26

Form: C310001-2 Rev A Issue date: 05/10/2024