

*Certificate of Calibration*

No. 0001076

TEMPERATURE	AS FOUND			LIMIT*		AS LEFT		
	Measured Value	In Tolerance	Out of Tolerance	Low	Hi	Measured Value	Fail	Pass
Curie Temperature ALUMEL® °C	161.89	<input type="checkbox"/>	<input checked="" type="checkbox"/>	148.9	159.5	151.39	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Curie Temperature Nickel °C	375.84	<input type="checkbox"/>	<input checked="" type="checkbox"/>	351.6	362.2	356.58	<input type="checkbox"/>	<input checked="" type="checkbox"/>

WEIGHT	AS FOUND			LIMIT*		AS LEFT		
	Measured Value	In Tolerance	Out of Tolerance	Low	Hi	Measured Value	Fail	Pass
Mass 1000 mg	1004.00	<input type="checkbox"/>	<input checked="" type="checkbox"/>	999.00	1001.00	999.6	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Mass 100 mg	100.50	<input type="checkbox"/>	<input checked="" type="checkbox"/>	99.90	100.10	100.00	<input type="checkbox"/>	<input checked="" type="checkbox"/>

\*(r) Repeatability values as determined by ASTM methods.

Temperature Calibration: ASTM E1582  
Mass Scale Calibration: ASTM E2040

ALUMEL® is a registered trademark of Concept Alloys, Whitmore Lake, Michigan.

